

C E R T I F I C A T E
of Conformity



Registration No.: AK 60021565 0001

Report No.: 28101799 001

Holder: Tecnocontrol Srl
Via Miglioli 47
20090 Segrate MI
Italia

Product: Analysis Equipment
(Combustion Product Analyser)

Identification: Central Unit Mod.: AN750
Probes Mod.: TS345, TS346, TS236, AN 510, TS352
The system is composed by:
- Central Unit AN750
- Probes TS345, TS346 (Tamb, Tcomb, Tsmoke), TS236 (O2),
AN 510 (electrochemical cell CO sensor, NO, SO2),
TS352 (Pcombustion chamber, Pchimney)
Specific Characteristics:
- Electro-chemical cell type CO Sensor

Tested acc. to: DLGS 3 April 2006, n° 152 part V, title I and II/04.06

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Cologne, 14.05.2008

Certification Body



TÜV Rheinland Product Safety GmbH - Am Grauen Stein - D-51105 Köln